

SAMPLE

検査成績表
<TEST REPORT>

1/1

品名<Model Name> (カルキュハイテスタ<CALCU HiTESTER>)
 形名<Model Number> (3208)
 製造番号<Serial No.> (No. 110712345)
 検査年月日<Test Date> (2011-12-31)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (23.0 °C, 50 %rh)
 検査周波数<Test Frequency> (55Hz)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>
確度<Accuracy>				
DCV	200 mV	180.0 mV	178.4 mV	~ 181.6 mV (180.0 mV)
	2 V	1.800 V	1.784 V	~ 1.816 V (1.800 V)
	20 V	18.00 V	17.84 V	~ 18.16 V (18.00 V)
	200 V	180.0 V	178.4 V	~ 181.6 V (180.0 V)
	1000 V	1000 V	989 V	~ 1011 V (1000 V)
ACV	2 V	1.800 V	1.776 V	~ 1.824 V (1.800 V)
	20 V	18.00 V	17.81 V	~ 18.19 V (18.00 V)
	200 V	180.0 V	178.1 V	~ 181.9 V (180.0 V)
	600 V	600 V	591 V	~ 609 V (600 V)
DCmA	20 mA	18.00 mA	17.77 mA	~ 18.23 mA (18.00 mA)
	200 mA	180.0 mA	177.7 mA	~ 182.3 mA (180.0 mA)
ACmA	20 mA	18.00 mA	17.72 mA	~ 18.28 mA (18.00 mA)
	200 mA	180.0 mA	177.2 mA	~ 182.8 mA (180.0 mA)
Ω	200 Ω	180.0 Ω	178.6 Ω	~ 181.4 Ω (180.0 Ω)
	2 k Ω	1.800 k Ω	1.786 k Ω	~ 1.814 k Ω (1.800 k Ω)
	20 k Ω	18.00 k Ω	17.86 k Ω	~ 18.14 k Ω (18.00 k Ω)
	200 k Ω	180.0 k Ω	178.6 k Ω	~ 181.4 k Ω (180.0 k Ω)
	2000 k Ω	1800 k Ω	1777 k Ω	~ 1823 k Ω (1800 k Ω)
LP Ω	2 k Ω	1.800 k Ω	1.778 k Ω	~ 1.822 k Ω (1.800 k Ω)
	20 k Ω	18.00 k Ω	17.78 k Ω	~ 18.22 k Ω (18.00 k Ω)
	200 k Ω	180.0 k Ω	177.8 k Ω	~ 182.2 k Ω (180.0 k Ω)
	2000 k Ω	1800 k Ω	1769 k Ω	~ 1831 k Ω (1800 k Ω)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()