

SAMPLE
 検査成績表
 <INSPECTION DATA SHEET>

品名<Model Name> (電圧ロガー<VOLTAGE LOGGER>)
 形名<Model Number> (LR5042)
 製造番号<Serial No.> (No. 110712345)
 検査年月日<Inspection Date> (2011-07-01)
 (<YYYY-MM-DD>)
 検査条件<Inspection Condition> (23.0 °C, 50 %rh)

項目 <Item>	レンジ <Range>	検査点 <Inspection Point>	許容範囲 <Tolerance>	表示値 <Indicated Value>
--------------	----------------	---------------------------	---------------------	--------------------------

直流電圧精度<DC Voltage Accuracy>

5 V	0.000 V	-0.005 V ~ 0.005 V	(0.000 V)
	4.000 V	3.975 V ~ 4.025 V	(4.000 V)

総合判定<Judgment> (PASS)	検査者<Inspected by> ()	承認者<Approved by> ()
----------------------------	--------------------------	-------------------------