

SAMPLE
 検査成績表  
 <INSPECTION DATA SHEET>

品名<Model Name> ( 電圧ロガー<VOLTAGE LOGGER> )  
 形名<Model Number> ( LR5043 )  
 製造番号<Serial No.> ( No. 110712345 )  
 検査年月日<Inspection Date> ( 2011-07-01 )  
 ( <YYYY-MM-DD> )  
 検査条件<Inspection Condition> ( 23.0 °C, 50 %rh )

項目 <Item>	レンジ <Range>	検査点 <Inspection Point>	許容範囲 <Tolerance>	表示値 <Indicated Value>
--------------	----------------	---------------------------	---------------------	--------------------------

直流電圧精度<DC Voltage Accuracy>

50 V	0.00 V	-0.05 V ~ 0.05 V	( 0.00 V )
	40.00 V	39.75 V ~ 40.25 V	( 40.00 V )

総合判定<Judgment> ( PASS )	検査者<Inspected by> ( )	承認者<Approved by> ( )
----------------------------	--------------------------	-------------------------