

**SAMPLE**

検査成績表

<TEST REPORT>

品名<Model Name> ( パルス発生ユニット<PULSE GENERATOR UNIT> )  
 形名<Model Number> ( MR8791 )  
 製造番号<Serial No.> (No. 190912345 )  
 検査年月日<Test Date> ( 2019-09-19 )  
 <YYYY-MM-DD>  
 検査条件<Test Conditions>( 25 °C, 50 %rh )

項目 <Item>	設定 <Set Value>	結果 <Result>
--------------	-------------------	----------------

セルフチェック <Self Check>

出力<Output>  
L:ロジック<Logic> / OC:オープンコレクタ<Open Collector> ch 8 7 6 5 4 3 2 1

(1) L	(0x55)	0 1 0 1 0 1 0 1	( PASS )
(2) L	(0xAA)	1 0 1 0 1 0 1 0	( PASS )
(3) OC	(0x55)	0 1 0 1 0 1 0 1	( PASS )
(4) OC	(0xAA)	1 0 1 0 1 0 1 0	( PASS )

接続確認<Connect Check>

出力<Output>  
L:ロジック<Logic> / OC:オープンコレクタ<Open Collector>

(1) L	パルス<Pulse>	1ch	20,000.0 Hz	duty	2.0% ("H")	( PASS )
(2) L	パルス<Pulse>	2ch	20,000.0 Hz	duty	2.1% ("H")	( PASS )
(3) L	パルス<Pulse>	3ch	20,000.0 Hz	duty	50.0% ("H")	( PASS )
(4) L	パルス<Pulse>	4ch	20,000.0 Hz	duty	98.0% ("H")	( PASS )
(5) L	パルス<Pulse>	5ch	1,000.0 Hz	duty	0.1% ("H")	( PASS )
(6) L	パルス<Pulse>	6ch	1,000.0 Hz	duty	0.2% ("H")	( PASS )
(7) L	パルス<Pulse>	7ch	1,000.0 Hz	duty	50.0% ("H")	( PASS )
(8) L	パルス<Pulse>	8ch	1,000.0 Hz	duty	99.9% ("H")	( PASS )
(9) OC	パルス<Pulse>	1ch	20,000.0 Hz	duty	10.0% ("H")	( PASS )
(10) OC	パルス<Pulse>	2ch	20,000.0 Hz	duty	20.0% ("H")	( PASS )
(11) OC	パルス<Pulse>	3ch	20,000.0 Hz	duty	30.0% ("H")	( PASS )
(12) OC	パルス<Pulse>	4ch	20,000.0 Hz	duty	50.0% ("H")	( PASS )
(13) OC	パルス<Pulse>	5ch	20,000.0 Hz	duty	60.0% ("H")	( PASS )
(14) OC	パルス<Pulse>	6ch	20,000.0 Hz	duty	70.0% ("H")	( PASS )
(15) OC	パルス<Pulse>	7ch	20,000.0 Hz	duty	80.0% ("H")	( PASS )
(16) OC	パルス<Pulse>	8ch	20,000.0 Hz	duty	90.0% ("H")	( PASS )
(17) L	パターン<Pattern>	1ch	120,000.0 bps	1-0繰返しパターン<Cyclic Pattern>		( PASS )
(18) L	パターン<Pattern>	2ch	120,000.0 bps	1-0繰返しパターン<Cyclic Pattern>		( PASS )
(19) L	パターン<Pattern>	3ch	120,000.0 bps	1-0繰返しパターン<Cyclic Pattern>		( PASS )
(20) L	パターン<Pattern>	4ch	120,000.0 bps	1-0繰返しパターン<Cyclic Pattern>		( PASS )
(21) OC	パターン<Pattern>	5ch	120,000.0 bps	1-0繰返しパターン<Cyclic Pattern>		( PASS )
(22) OC	パターン<Pattern>	6ch	120,000.0 bps	1-0繰返しパターン<Cyclic Pattern>		( PASS )
(23) OC	パターン<Pattern>	7ch	120,000.0 bps	1-0繰返しパターン<Cyclic Pattern>		( PASS )
(24) OC	パターン<Pattern>	8ch	120,000.0 bps	1-0繰返しパターン<Cyclic Pattern>		( PASS )

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
( PASS )	( )	( )