

SAMPLE

# 検査成績表

<TEST REPORT>

1/1

|                       |       |                  |   |
|-----------------------|-------|------------------|---|
| 品名<Model Name>        | (     | 電源<POWER SUPPLY> | ) |
| 形名<Model Number>      | (     | 3269             | ) |
| 製造番号<Serial No.>      | ( No. | 200412345        | ) |
| 検査年月日<Test Date>      | (     | 2020-04-19       | ) |
|                       |       | <YYYY-MM-DD>     |   |
| 検査条件<Test Conditions> | (     | 23.0 °C, 50 %rh  | ) |

| 項目<br><Item>                               | 電源チャネル<br><Power Supply Receptacles>              | 許容範囲<br><Tolerance>                              | 出力値<br><Indicated Value> |             |
|--|---|--|--------------------------|-------------|
| 出力電圧(無負荷)<br><Output Voltage<br>(No-Load)> | PROBE 1   | 11.5 V ~ 12.5 V                                  | ( 12.0 V )               |             |
|  |   | -12.5 V ~ -11.5 V                                | ( -12.0 V )              |             |
|  | PROBE 2   | 11.5 V ~ 12.5 V                                  | ( 12.0 V )               |             |
|  |   | -12.5 V ~ -11.5 V                                | ( -12.0 V )              |             |
|  | PROBE 3   | 11.5 V ~ 12.5 V                                  | ( 12.0 V )               |             |
|  |   | -12.5 V ~ -11.5 V                                | ( -12.0 V )              |             |
|  | PROBE 4   | 11.5 V ~ 12.5 V                                  | ( 12.0 V )               |             |
|  |   | -12.5 V ~ -11.5 V                                | ( -12.0 V )              |             |
|  | 出力電圧(負荷±2.5A)<br><Output Voltage<br>(Load ±2.5A)> | PROBE 1,PROBE 2<br>並列接続<br><Multiple Connection> | 11.5 V ~ 12.5 V          | ( 12.0 V )  |
|  |   |  | -12.5 V ~ -11.5 V        | ( -12.0 V ) |
|  |   |  |                          |             |

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

|                      |                   |                  |
|----------------------|-------------------|------------------|
| 総合判定<Overall Result> | 検査者<Inspected by> | 承認者<Approved by> |
| ( PASS )             | ( )               | ( )              |