

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name>	(パワーハイテスタ<POWER HiTESTER>)	
形名<Model Number>	(3333)	
製造番号<Serial No.>	(No. 160811700)	
検査年月日<Test Date>	(2016-08-04)	
	<YYYY-MM-DD>	
検査条件<Test Conditions>	(23.5 °C, 57 %rh)	
検査周波数<Test Frequency>	(55 Hz)	

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>	
確度<Accuracy>					
電圧 <U>	200V	0V	0.0 V ~ 0.0 V	(0.0 V)	
		20V	19.8 V ~ 20.2 V	(20.0 V)	
		100V	99.7 V ~ 100.3 V	(100.0 V)	
		200V	199.6 V ~ 200.4 V	(200.0 V)	
		300V	299.4 V ~ 300.6 V	(300.1 V)	
電流 <I>	50mA	0A	0.00 mA ~ 0.00 mA	(0.00 mA)	
		50mA	49.90 mA ~ 50.10 mA	(49.98 mA)	
		200mA	0.0 mA ~ 0.0 mA	(0.0 mA)	
	200mA	200mA	199.6 mA ~ 200.4 mA	(200.0 mA)	
		500mA	0.0 mA ~ 0.0 mA	(0.0 mA)	
		500mA	499.0 mA ~ 501.0 mA	(500.0 mA)	
	2A	0A	0.000 A ~ 0.000 A	(0.000 A)	
		0.2A	0.198 A ~ 0.202 A	(0.200 A)	
		1A	0.997 A ~ 1.003 A	(1.000 A)	
		2A	1.996 A ~ 2.004 A	(2.000 A)	
		3A	2.994 A ~ 3.006 A	(3.000 A)	
		5A	0.000 A ~ 0.000 A	(0.000 A)	
	5A	5A	4.990 A ~ 5.010 A	(4.999 A)	
		20A	0A	0.00 A ~ 0.00 A	(0.00 A)
			20A	19.96 A ~ 20.04 A	(20.00 A)
有効電力 <P> (cos φ=1)	200V 2A	0V 0A 0W	-0.4 W ~ 0.4 W	(0.0 W)	
		20V 0.2A 4W	3.6 W ~ 4.4 W	(4.0 W)	
		200V 0.2A 40W	39.6 W ~ 40.4 W	(40.0 W)	
		200V 1A 200W	199.4 W ~ 200.6 W	(200.1 W)	
		200V 2A 400W	399.2 W ~ 400.8 W	(400.1 W)	
		300V 3A 900W	898.2 W ~ 901.8 W	(900.0 W)	
(cos φ=0.5) LEAD LAG	200V 2A	200V 2A 200W	198.6 W ~ 201.4 W	(199.9 W)	
		200V 2A 200W	198.6 W ~ 201.4 W	(200.1 W)	

項目 <Item>	出力項目 <Output Item>	出力基底値 <Fiducial Value>	許容範囲 <Tolerance>	出力値 <Output>
D/A出力 <D/A Output>	U	0.2V	0.1900 V ~ 0.2100 V	(0.2007 V)
		2V	1.9900 V ~ 2.0100 V	(1.9999 V)
	I	0.2V	0.1900 V ~ 0.2100 V	(0.1996 V)
		2V	1.9900 V ~ 2.0100 V	(1.9998 V)
	P	0.2V	0.1900 V ~ 0.2100 V	(0.2002 V)
		2V	1.9900 V ~ 2.0100 V	(1.9997 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定(1年確度) <Overall Result>(Accuracy Tolerance After 1 Year) (PASS)	検査者<Inspected by> ()	承認者<Approved by> ()
------------------------------------------------------------------------------------	---------------------------------------	--------------------------------------

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name>	(パワーハイテスタ<POWER HiTESTER>)	
形名<Model Number>	(3333)	
製造番号<Serial No.>	(No. 160811700)	
検査年月日<Test Date>	(2016-08-04)	
	<YYYY-MM-DD>	
検査条件<Test Conditions>	(23.5 °C, 57 %rh)	
検査周波数<Test Frequency>	(55 Hz)	

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	表示値 <Indicated Value>	
確度<Accuracy>					
電圧 <U>	200V	0V	0.0 V ~ 0.0 V	(0.0 V)	
		20V	19.6 V ~ 20.4 V	(20.0 V)	
		100V	99.5 V ~ 100.5 V	(100.0 V)	
		200V	199.4 V ~ 200.6 V	(200.0 V)	
		300V	299.1 V ~ 300.9 V	(300.1 V)	
	電流 <I>	50mA	0A	0.00 mA ~ 0.00 mA	(0.00 mA)
			50mA	49.85 mA ~ 50.15 mA	(49.98 mA)
		200mA	0A	0.0 mA ~ 0.0 mA	(0.0 mA)
			200mA	199.4 mA ~ 200.6 mA	(200.0 mA)
		500mA	0A	0.0 mA ~ 0.0 mA	(0.0 mA)
500mA			498.5 mA ~ 501.5 mA	(500.0 mA)	
2A		0A	0.000 A ~ 0.000 A	(0.000 A)	
		0.2A	0.196 A ~ 0.204 A	(0.200 A)	
		1A	0.995 A ~ 1.005 A	(1.000 A)	
		2A	1.994 A ~ 2.006 A	(2.000 A)	
	3A	2.991 A ~ 3.009 A	(3.000 A)		
5A	0A	0.000 A ~ 0.000 A	(0.000 A)		
	5A	4.985 A ~ 5.015 A	(4.999 A)		
20A	0A	0.00 A ~ 0.00 A	(0.00 A)		
	20A	19.94 A ~ 20.06 A	(20.00 A)		
有効電力 <P> (cos φ=1)	200V 2A	0V 0A 0W	-0.8 W ~ 0.8 W	(0.0 W)	
		20V 0.2A 4W	3.2 W ~ 4.8 W	(4.0 W)	
		200V 0.2A 40W	39.2 W ~ 40.8 W	(40.0 W)	
		200V 1A 200W	199.0 W ~ 201.0 W	(200.1 W)	
		200V 2A 400W	398.8 W ~ 401.2 W	(400.1 W)	
		300V 3A 900W	897.3 W ~ 902.7 W	(900.0 W)	
		200V 2A 200W	198.2 W ~ 201.8 W	(199.9 W)	
(cos φ=0.5) LEAD LAG	200V 2A	200V 2A 200W	198.2 W ~ 201.8 W	(200.1 W)	
		200V 2A 200W	198.2 W ~ 201.8 W	(200.1 W)	

項目 <Item>	出力項目 <Output Item>	出力基底値 <Fiducial value>	許容範囲 <Tolerance>	出力値 <Output>
D/A出力 <D/A Output>	U	0.2V	0.1900 V ~ 0.2100 V	(0.2007 V)
		2V	1.9900 V ~ 2.0100 V	(1.9999 V)
	I	0.2V	0.1900 V ~ 0.2100 V	(0.1996 V)
		2V	1.9900 V ~ 2.0100 V	(1.9998 V)
	P	0.2V	0.1900 V ~ 0.2100 V	(0.2002 V)
		2V	1.9900 V ~ 2.0100 V	(1.9997 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定(3年確度) <Overall Result>(Accuracy Tolerance After 3 Year) (PASS)	検査者<Inspected by> ()	承認者<Approved by> ()
------------------------------------------------------------------------------------	---------------------------------------	--------------------------------------