

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E. E. CORPORATION

1/1

品名<Model Name> (アナログメグオームハイテスタ<ANALOG MΩ HiTESTER>)
 形名<Model Number> (3490)
 製造番号<Serial No.> (No.170765432)
 検査年月日<Test Date> (2017-07-12)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 56 %rh)

| 項目 <Item> | レンジ <Range> | 計器指示値 <Indicated Value> | 許容範囲 <Tolerance> | 校正値 <Calibration Value> |
|--|----------------|----------------------------|-----------------------|----------------------------|
| 確度(固有誤差) <Accuracy (Intrinsic error)> | | | | |
| 250V | | 0.01 MΩ | 0.0070 MΩ ~ 0.0137 MΩ | (0.0100 MΩ) |
| | | 0.05 MΩ | 0.0398 MΩ ~ 0.0624 MΩ | (0.0500 MΩ) |
| | | 1 MΩ | 0.820 MΩ ~ 1.219 MΩ | (1.000 MΩ) |
| | | 50 MΩ | 41.1 MΩ ~ 60.9 MΩ | (50.0 MΩ) |
| | | 100 MΩ | 82.1 MΩ ~ 121.8 MΩ | (100.0 MΩ) |
| 500V | | 0.01 MΩ | 0.0070 MΩ ~ 0.0137 MΩ | (0.0100 MΩ) |
| | | 0.05 MΩ | 0.0398 MΩ ~ 0.0624 MΩ | (0.0500 MΩ) |
| | | 1 MΩ | 0.820 MΩ ~ 1.219 MΩ | (1.000 MΩ) |
| | | 50 MΩ | 41.1 MΩ ~ 60.9 MΩ | (50.0 MΩ) |
| | | 100 MΩ | 82.1 MΩ ~ 121.8 MΩ | (100.0 MΩ) |
| 1000V | | 0.5 MΩ | 0.410 MΩ ~ 0.610 MΩ | (0.500 MΩ) |
| | | 2 MΩ | 1.641 MΩ ~ 2.438 MΩ | (2.000 MΩ) |
| | | 50 MΩ | 41.1 MΩ ~ 60.9 MΩ | (50.0 MΩ) |
| | | 1000 MΩ | 821 MΩ ~ 1218 MΩ | (1000 MΩ) |
| | | 4000 MΩ | 3284 MΩ ~ 4873 MΩ | (4000 MΩ) |
| 3Ω | 2.50 Ω | 2.410 Ω ~ 2.590 Ω | (2.500 Ω) | |
| 30Ω | 25.00 Ω | 24.10 Ω ~ 25.90 Ω | (25.00 Ω) | |
| ACV (55Hz) | 100 V | 70 V ~ 130 V | (100 V) | |
| | 500 V | 470 V ~ 530 V | (500 V) | |

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

| | | |
|----------------------------------|--------------------------|-------------------------|
| 総合判定<Overall Result> (PASS) | 検査者<Inspected by> () | 承認者<Approved by> () |
|----------------------------------|--------------------------|-------------------------|