

SAMPLE 検査成績表  
<TEST REPORT>

品名<Model Name> ( ストレインユニット<STRAIN UNIT> )  
形名<Model Number> ( U8969 )  
製造番号<Serial No.> ( No. 170919681 )  
検査年月日<Test Date> ( 2017-09-08 )  
( <YYYY-MM-DD> )  
検査条件<Test Conditions> ( 25.0 °C, 57 %rh )

項目 <Item>	レンジ <Range>	検査点 <Test Point>	許容範囲 <Tolerance>	測定値 <Measured Value>
--------------	----------------	---------------------	---------------------	-------------------------

1. 測定精度<Measurement Accuracy>

				CH1	CH2
20 $\mu\epsilon$ /DIV	0 $\mu\epsilon$	-6.00 $\mu\epsilon$ ~ 6.00 $\mu\epsilon$	( 0.05 $\mu\epsilon$ )	( -0.13 $\mu\epsilon$ )	
	400 $\mu\epsilon$	394.00 $\mu\epsilon$ ~ 406.00 $\mu\epsilon$	( 400.05 $\mu\epsilon$ )	( 400.24 $\mu\epsilon$ )	
50 $\mu\epsilon$ /DIV	0 $\mu\epsilon$	-9.00 $\mu\epsilon$ ~ 9.00 $\mu\epsilon$	( 0.24 $\mu\epsilon$ )	( -0.20 $\mu\epsilon$ )	
	1000 $\mu\epsilon$	991.00 $\mu\epsilon$ ~ 1009.00 $\mu\epsilon$	( 1000.08 $\mu\epsilon$ )	( 999.68 $\mu\epsilon$ )	
100 $\mu\epsilon$ /DIV	0 $\mu\epsilon$	-14.00 $\mu\epsilon$ ~ 14.00 $\mu\epsilon$	( -0.16 $\mu\epsilon$ )	( -0.24 $\mu\epsilon$ )	
	2000 $\mu\epsilon$	1986.00 $\mu\epsilon$ ~ 2014.00 $\mu\epsilon$	( 1999.60 $\mu\epsilon$ )	( 1999.76 $\mu\epsilon$ )	
200 $\mu\epsilon$ /DIV	0 $\mu\epsilon$	-24.0 $\mu\epsilon$ ~ 24.0 $\mu\epsilon$	( 0.5 $\mu\epsilon$ )	( -0.3 $\mu\epsilon$ )	
	4000 $\mu\epsilon$	3976.0 $\mu\epsilon$ ~ 4024.0 $\mu\epsilon$	( 3999.4 $\mu\epsilon$ )	( 3998.7 $\mu\epsilon$ )	
500 $\mu\epsilon$ /DIV	0 $\mu\epsilon$	-54.0 $\mu\epsilon$ ~ 54.0 $\mu\epsilon$	( 0.4 $\mu\epsilon$ )	( -1.2 $\mu\epsilon$ )	
	10000 $\mu\epsilon$	9946.0 $\mu\epsilon$ ~ 10054.0 $\mu\epsilon$	( 9998.0 $\mu\epsilon$ )	( 9998.0 $\mu\epsilon$ )	
1000 $\mu\epsilon$ /DIV	0 $\mu\epsilon$	-104.0 $\mu\epsilon$ ~ 104.0 $\mu\epsilon$	( -0.8 $\mu\epsilon$ )	( -1.6 $\mu\epsilon$ )	
	20000 $\mu\epsilon$	19896.0 $\mu\epsilon$ ~ 20104.0 $\mu\epsilon$	( 19996.8 $\mu\epsilon$ )	( 19995.2 $\mu\epsilon$ )	

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result> ( PASS )	検査者<Inspected by> ( )	承認者<Approved by> ( )
----------------------------------	--------------------------	-------------------------