

SAMPLE

# 検査成績表

<TEST REPORT>

品名<Model Name>	( デジタルボルトメータユニット<DIGITAL VOLTMETER UNIT> )	( )
形名<Model Number>	( U8991 )	( )
製造番号<Serial No.>	( No.180711111 )	( )
検査年月日<Test Date>	( 2018-07-25 )	( )
	<YYYY-MM-DD>	
検査条件<Test Conditions>	( 24.4 °C, 56 %rh )	( )

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 <Calibration Value>
1.測定精度<Measurement Accuracy>				
	1 V f.s.	0 V	-0.000025 V ~ 0.000025 V	CH1 ( 0.000000 V ) CH2 ( 0.000002 V ) CH3 ( 0.000001 V ) CH4 ( 0.000001 V )
		0.5 V	0.499875 V ~ 0.500125 V	CH1 ( 0.500001 V ) CH2 ( 0.499999 V ) CH3 ( 0.500000 V ) CH4 ( 0.500001 V )
		-0.5 V	-0.500125 V ~ -0.499875 V	CH1 ( -0.499999 V ) CH2 ( -0.499998 V ) CH3 ( -0.499998 V ) CH4 ( -0.499998 V )
		1 V	0.999775 V ~ 1.000225 V	CH1 ( 1.000002 V ) CH2 ( 0.999999 V ) CH3 ( 0.999999 V ) CH4 ( 1.000001 V )
		-1 V	-1.000225 V ~ -0.999775 V	CH1 ( -1.000002 V ) CH2 ( -1.000000 V ) CH3 ( -1.000001 V ) CH4 ( -1.000001 V )
	10 V f.s.	0 V	-0.000025 V ~ 0.000025 V	CH1 ( -0.000001 V ) CH2 ( -0.000001 V ) CH3 ( -0.000001 V ) CH4 ( 0.000000 V )
		5 V	4.99875 V ~ 5.00125 V	CH1 ( 5.000001 V ) CH2 ( 4.999999 V ) CH3 ( 4.999999 V ) CH4 ( 5.000001 V )
		-5 V	-5.00125 V ~ -4.99875 V	CH1 ( -4.999999 V ) CH2 ( -4.999998 V ) CH3 ( -4.999999 V ) CH4 ( -4.999999 V )
		10 V	9.99775 V ~ 10.00225 V	CH1 ( 10.000002 V ) CH2 ( 9.999999 V ) CH3 ( 10.000000 V ) CH4 ( 10.000002 V )
		-10 V	-10.00225 V ~ -9.99775 V	CH1 ( -10.000002 V ) CH2 ( -9.999999 V ) CH3 ( -10.000001 V ) CH4 ( -10.000002 V )

備考<Note> FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
( PASS )	( )	( )

# SAMPLE

## 検査成績表

<TEST REPORT>

**HIOKI**  
HIOKI E. E. CORPORATION

2/2

品名<Model Name>	( デジタルボルトメータユニット<DIGITAL VOLTMETER UNIT> )	
形名<Model Number>	( U8991 )	
製造番号<Serial No.>	( No.180711111 )	
検査年月日<Test Date>	( 2018-07-25 )	
	<YYYY-MM-DD>	

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	校正値 <Calibration Value>
<b>1.測定精度&lt;Measurement Accuracy&gt;</b>				
	100 V f.s.	0 V	-0.0025 V ~ 0.0025 V	CH1( 0.0000 V) CH2( -0.0001 V) CH3( -0.0001 V) CH4( 0.0000 V)
		50 V	49.9875 V ~ 50.0125 V	CH1( 50.0001 V) CH2( 49.9999 V) CH3( 50.0001 V) CH4( 50.0002 V)
		-50 V	-50.0125 V ~ -49.9875 V	CH1( -49.9999 V) CH2( -49.9998 V) CH3( -49.9999 V) CH4( -49.9999 V)
		100 V	99.9775 V ~ 100.0225 V	CH1( 100.0002 V) CH2( 99.9999 V) CH3( 100.0001 V) CH4( 100.0002 V)
		-100 V	-100.0225 V ~ -99.9775 V	CH1( -100.0003 V) CH2( -99.9999 V) CH3( -100.0001 V) CH4( -100.0002 V)