

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name> (クランプ接地抵抗計)
<CLAMP ON EARTH TESTER>
形名<Model Number> (FT6380)
製造番号<Serial No.> (No. 130912345)
検査年月日<Test Date> (2013-09-01)
<YYYY-MM-DD>
検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	*1 表示値 <Indicated Value>
抵抗確度<Resistance Accuracy>				
*2 1600 / 1200 Ω		1200 Ω	960 Ω ~ 1440 Ω	(1220 Ω)
*2 1200 / 600 Ω		600 Ω	530 Ω ~ 670 Ω	(600 Ω)
*2 600 / 400 Ω		400 Ω	375 Ω ~ 425 Ω	(404 Ω)
200.0 Ω		170.0 Ω	163.9 Ω ~ 176.1 Ω	(170.4 Ω)
*2 200.0 / 100.0 Ω		100.0 Ω	98.0 Ω ~ 102.0 Ω	(100.0 Ω)
*2 100.0 / 50.0 Ω		50.0 Ω	49.2 Ω ~ 50.8 Ω	(50.0 Ω)
20.00 Ω		15.00 Ω	14.73 Ω ~ 15.27 Ω	(14.99 Ω)
2.00 Ω		1.50 Ω	1.46 Ω ~ 1.54 Ω	(1.49 Ω)
0.20 Ω		0.15 Ω	0.13 Ω ~ 0.17 Ω	(0.15 Ω)

備考<Note>

*1. FAIL判定箇所は、グレー表示としています。

<FAIL decision points are highlighted in gray.>

*2. ハードウェアレンジが共通のため同一ポイントで検査を行っています。

判定は、より高確度のレンジの確度を適用しています。

<Due to the common hardware ranges, the tests are made on the same points.

Judgment is done by using higher accuracy range.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E.E. CORPORATION

2/2

製造番号<Serial No.> (No. 130912345)
検査年月日<Test Date> (2013-09-01)
<YYYY-MM-DD>

項目 <Item>	レンジ／フィルタ <Range/Filter>	入力／周波数 <Input/Frequency>	許容範囲 <Tolerance>	*1 表示値 <Indicated Value>
電流確度<Current Accuracy>				
	60.0 A	OFF	50.0 A 55 Hz	48.5 A ~ 51.5 A (50.0 A)
	20.00 A	OFF	17.00 A 55 Hz	16.61 A ~ 17.39 A (17.00 A)
	2.000 A	OFF	1.700 A 55 Hz	1.661 A ~ 1.739 A (1.700 A)
	200.0 mA	OFF	170.0 mA 55 Hz	166.1 mA ~ 173.9 mA (170.0 mA)
		OFF	100.0 mA 55 Hz	97.5 mA ~ 102.5 mA (100.0 mA)
		OFF	30.0 mA 55 Hz	28.9 mA ~ 31.1 mA (30.0 mA)
	20.00 mA	OFF	17.00 mA 55 Hz	16.61 mA ~ 17.39 mA (17.00 mA)
		OFF	1.70 mA 55 Hz	1.62 mA ~ 1.78 mA (1.70 mA)
		OFF	0.00 mA -	-0.05 mA ~ 0.05 mA (0.00 mA)
	60.0 A	ON	50.0 A 55 Hz	48.5 A ~ 51.5 A (50.0 A)
	20.00 A	ON	17.00 A 55 Hz	16.61 A ~ 17.39 A (16.98 A)
	2.000 A	ON	1.700 A 55 Hz	1.661 A ~ 1.739 A (1.700 A)
	200.0 mA	ON	170.0 mA 55 Hz	166.1 mA ~ 173.9 mA (170.0 mA)
	20.00 mA	ON	17.00 mA 55 Hz	16.61 mA ~ 17.39 mA (17.00 mA)
		ON	1.70 mA 55 Hz	1.62 mA ~ 1.78 mA (1.70 mA)
		ON	0.00 mA -	-0.05 mA ~ 0.05 mA (0.00 mA)
周波数特性 <Frequency Characteristic>	200.0 mA	OFF	170.0 mA 30 Hz	165.3 mA ~ 174.7 mA (169.4 mA)
		OFF	170.0 mA 400 Hz	165.3 mA ~ 174.7 mA (170.0 mA)
	200.0 mA	ON	170.0 mA 45 Hz	166.1 mA ~ 173.9 mA (169.7 mA)
		ON	170.0 mA 66 Hz	166.1 mA ~ 173.9 mA (169.8 mA)