

SAMPLE

検査成績表
<TEST REPORT>HIOKI
HIOKI E. E. CORPORATION

1/1

品名<Model Name> (ストレインユニット<STRAIN UNIT>)
形名<Model Number> (8960)
製造番号<Serial No.> (No. 100702950)
検査年月日<Test Date> (2010-07-01)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (25.0 °C, 69 %rh)

項目 <Item>	レンジ <Range>	検査点 <Test Point>	許容範囲 <Tolerance>	測定値 <Measured Value>
--------------	----------------	---------------------	---------------------	-------------------------

1.DC振幅確度<DC Amplitude Accuracy>

ブリッジ電圧 2V <Bridge Voltage 2V>

CH1

CH2

20 $\mu\epsilon$ /DIV	400 $\mu\epsilon$	396.400 $\mu\epsilon$ ~ 403.600 $\mu\epsilon$ (399.725 $\mu\epsilon$) (400.275 $\mu\epsilon$)
50 $\mu\epsilon$ /DIV	1000 $\mu\epsilon$	994.000 $\mu\epsilon$ ~ 1006.000 $\mu\epsilon$ (1000.094 $\mu\epsilon$) (1000.281 $\mu\epsilon$)
100 $\mu\epsilon$ /DIV	2000 $\mu\epsilon$	1990.000 $\mu\epsilon$ ~ 2010.000 $\mu\epsilon$ (2000.125 $\mu\epsilon$) (2000.812 $\mu\epsilon$)
200 $\mu\epsilon$ /DIV	4000 $\mu\epsilon$	3982.00 $\mu\epsilon$ ~ 4018.00 $\mu\epsilon$ (4000.12 $\mu\epsilon$) (4000.50 $\mu\epsilon$)
500 $\mu\epsilon$ /DIV	10000 $\mu\epsilon$	9958.00 $\mu\epsilon$ ~ 10042.00 $\mu\epsilon$ (9999.69 $\mu\epsilon$) (10001.56 $\mu\epsilon$)
1000 $\mu\epsilon$ /DIV	20000 $\mu\epsilon$	19918.00 $\mu\epsilon$ ~ 20082.00 $\mu\epsilon$ (20002.50 $\mu\epsilon$) (20008.12 $\mu\epsilon$)

ブリッジ電圧 5V <Bridge Voltage 5V>

20 $\mu\epsilon$ /DIV	400 $\mu\epsilon$	396.400 $\mu\epsilon$ ~ 403.600 $\mu\epsilon$ (400.187 $\mu\epsilon$) (400.062 $\mu\epsilon$)
-----------------------	-------------------	---

ブリッジ電圧 10V <Bridge Voltage 10V>

20 $\mu\epsilon$ /DIV	400 $\mu\epsilon$	396.400 $\mu\epsilon$ ~ 403.600 $\mu\epsilon$ (400.062 $\mu\epsilon$) (400.038 $\mu\epsilon$)
-----------------------	-------------------	---

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()