

SAMPLE

検査成績表
<TEST REPORT>HIOKI
HIOKI E. E. CORPORATION

1/1

品名 <Model Name> (ラインディップディテクタ <LINE DIP DETECTOR>)
形名 <Model Number> (9308)
製造番号 <Serial No.> (No. 110801234)
検査年月日 <Test Date> (2011-08-05)
(<YYYY-MM-DD>)
検査条件 <Test Conditions> (25.5 °C, 55 %rh)

機能
<Function>

項目 <Item>	FREQ	レンジ <Range> DIP	LINE.V	検査点 <Test Point>	入力 (LED) <Input (LED)>	結果 <Result>
--------------	------	-----------------------	--------	---------------------	---------------------------	----------------

1.レンジ <Range>

50Hz	90%	AC	100V	AC 91V/50Hz	点灯 (ON)	(PASS)
50Hz	90%	AC	120V	AC 109V/50Hz	点灯 (ON)	(PASS)
50Hz	80%	AC	100V	AC 81V/50Hz	点灯 (ON)	(PASS)
50Hz	80%	AC	120V	AC 97V/50Hz	点灯 (ON)	(PASS)
50Hz	90%	AC	100V	AC 89V/50Hz	消灯 (OFF)	(PASS)
50Hz	90%	AC	120V	AC 107V/50Hz	消灯 (OFF)	(PASS)
50Hz	80%	AC	100V	AC 79V/50Hz	消灯 (OFF)	(PASS)
50Hz	80%	AC	120V	AC 95V/50Hz	消灯 (OFF)	(PASS)

2.LED <LED>

60Hz	90%	AC	100V	AC 100V/60Hz	点灯 (ON)	(PASS)
60Hz	90%	AC	90V	AC 90V/60Hz	消灯 (OFF)	(PASS)

備考<Note>

総合判定 <Overall Result>	検査者 <Inspected By>	承認者 <Approved By>
(PASS)	()	()