

SAMPLE

検査成績表

<TEST REPORT>

HIOKI
HIOKI E. E. CORPORATION

1/1

品名<Model Name> (アナログメグオームハイテスタ<ANALOG MΩ HiTESTER>)
 形名<Model Number> (IR4082) *1
 製造番号<Serial No.> (No.180598765)
 検査年月日<Test Date> (2018-05-09)
 (<YYYY-MM-DD>)
 検査条件<Test Conditions> (23.0 °C, 60 %rh)

項目 <Item>	レンジ <Range>	計器指示値 <Indicated Value>	許容範囲 <Tolerance>	校正値 <Calibration Value>
確度(許容差) <Accuracy (Limit Deviation Tolerance)>				
125V	0.01 MΩ	0.0090 MΩ	~ 0.0110 MΩ	(0.0100 MΩ)
	0.1 MΩ	0.0950 MΩ	~ 0.1050 MΩ	(0.1000 MΩ)
	2 MΩ	1.90 MΩ	~ 2.10 MΩ	(2.00 MΩ)
	50 MΩ	47.5 MΩ	~ 52.5 MΩ	(50.0 MΩ)
	100 MΩ	90 MΩ	~ 110 MΩ	(100 MΩ)
250V	0.01 MΩ	0.0090 MΩ	~ 0.0110 MΩ	(0.0100 MΩ)
	0.1 MΩ	0.0950 MΩ	~ 0.1050 MΩ	(0.1000 MΩ)
	2 MΩ	1.90 MΩ	~ 2.10 MΩ	(2.00 MΩ)
	50 MΩ	47.5 MΩ	~ 52.5 MΩ	(50.0 MΩ)
	100 MΩ	90 MΩ	~ 110 MΩ	(100 MΩ)
500V	0.01 MΩ	0.0090 MΩ	~ 0.0110 MΩ	(0.0100 MΩ)
	0.1 MΩ	0.0950 MΩ	~ 0.1050 MΩ	(0.1000 MΩ)
	2 MΩ	1.90 MΩ	~ 2.10 MΩ	(2.00 MΩ)
	50 MΩ	47.5 MΩ	~ 52.5 MΩ	(50.0 MΩ)
	100 MΩ	90 MΩ	~ 110 MΩ	(100 MΩ)
ACV (55Hz)	100 V	70 V	~ 130 V	(100 V)
	500 V	470 V	~ 530 V	(500 V)
Ω × 1	10.0 Ω	7.0 Ω	~ 13.0 Ω	(10.0 Ω)
	50.0 Ω	47.0 Ω	~ 53.0 Ω	(50.0 Ω)
	100.0 Ω	97.0 Ω	~ 103.0 Ω	(100.0 Ω)
Ω × 10	100 Ω	70 Ω	~ 130 Ω	(100 Ω)
	500 Ω	470 Ω	~ 530 Ω	(500 Ω)
	1000 Ω	970 Ω	~ 1030 Ω	(1000 Ω)

項目 <Item>	結果 <Result>
機能<Function>	
地電圧<Earth Voltage>	(PASS)

備考<Note>

- ・ FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>
- ・ *1 形名欄の記載形名は、製品本体表記に基づいています。

<The model number stated in this document is based on the model number indicated on the instrument.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()