

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> (アナログユニット<ANALOG UNIT>)
形名<Model Number> (MR8901)
製造番号<Serial No.> (No. 110812345)
検査年月日<Test Date> (2011-08-22)
<YYYY-MM-DD>
検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	レンジ <Range>	入力 <Input>	許容範囲 <Tolerance>	CH1	測定値<Measured Value>			
					CH2	CH3	CH4	

確度<Accuracy>

DC振幅

<DC Amplitude>

5 mV/DIV	0 mV	DC	-0.500 mV ~ 0.500 mV	(0.000 mV)	(0.000 mV)	(0.000 mV)	(0.000 mV)
5 mV/DIV	-100 mV	DC	-100.500 mV ~ -99.500 mV	(-100.000 mV)	(-100.000 mV)	(-100.000 mV)	(-100.000 mV)
5 mV/DIV	100 mV	DC	99.500 mV ~ 100.500 mV	(100.000 mV)	(100.000 mV)	(100.000 mV)	(100.000 mV)
10 mV/DIV	200 mV	DC	199.000 mV ~ 201.000 mV	(200.000 mV)	(200.000 mV)	(200.000 mV)	(200.000 mV)
20 mV/DIV	400 mV	DC	398.000 mV ~ 402.000 mV	(400.000 mV)	(400.000 mV)	(400.000 mV)	(400.000 mV)
50 mV/DIV	1 V	DC	0.99500 V ~ 1.00500 V	(1.00000 V)	(1.00000 V)	(1.00000 V)	(1.00000 V)
100 mV/DIV	2 V	DC	1.99000 V ~ 2.01000 V	(2.00000 V)	(2.00000 V)	(2.00000 V)	(2.00000 V)
200 mV/DIV	4 V	DC	3.98000 V ~ 4.02000 V	(4.00000 V)	(4.00000 V)	(4.00000 V)	(4.00000 V)
500 mV/DIV	10 V	DC	9.9500 V ~ 10.0500 V	(10.0000 V)	(10.0000 V)	(10.0000 V)	(10.0000 V)
1 V/DIV	20 V	DC	19.9000 V ~ 20.1000 V	(20.0000 V)	(20.0000 V)	(20.0000 V)	(20.0000 V)
2 V/DIV	40 V	DC	39.8000 V ~ 40.2000 V	(40.0000 V)	(40.0000 V)	(40.0000 V)	(40.0000 V)
5 V/DIV	100 V	DC	99.500 V ~ 100.500 V	(100.000 V)	(100.000 V)	(100.000 V)	(100.000 V)
10 V/DIV	200 V	DC	199.000 V ~ 201.000 V	(200.000 V)	(200.000 V)	(200.000 V)	(200.000 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected By>	承認者<Approved By>
(PASS)	()	()