

SAMPLE

検査成績表

<TEST REPORT>

1/1

品名<Model Name> ('アナログユニット<ANALOG UNIT>)
 形名<Model Number> (8942)
 製造番号<Serial No.> (No. 091234567)
 検査年月日<Test Date> (2011-11-03)
 <YYYY-MM-DD>
 検査条件<Test Conditions> (23.0 °C, 50 %rh)

項目 <Item>	レンジ <Range>	試験点 <Test Point>	許容範囲 <Tolerance>	測定結果 <Measured Value>
--------------	----------------	---------------------	---------------------	--------------------------

ゼロポジション確度<Zero Position>

5 mV/DIV 0 mV ± 0.400 mV (0.400 mV)

確度<Accuracy>

5 mV/DIV	40 mV	±	0.90 mV	(40.40 mV)
10 mV/DIV	80 mV	±	1.80 mV	(80.80 mV)
20 mV/DIV	160 mV	±	3.6 mV	(161.6 mV)
50 mV/DIV	400 mV	±	9.0 mV	(404.0 mV)
100 mV/DIV	800 mV	±	18 mV	(812 mV)
200 mV/DIV	1.6 V	±	0.036 V	(1.616 V)
500 mV/DIV	4 V	±	0.090 V	(4.040 V)
1 V/DIV	8 V	±	0.18 V	(8.08 V)
2 V/DIV	16 V	±	0.36 V	(16.16 V)
5 V/DIV	40 V	±	0.90 V	(40.40 V)
10 V/DIV	80 V	±	1.80 V	(80.80 V)
20 V/DIV	160 V	±	3.6 V	(161.6 V)

備考<Note>

FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

総合判定<Overall Result>	検査者<Inspected by>	承認者<Approved by>
(PASS)	()	()