

SAMPLE

検査成績表

<TEST REPORT>

品名<Model Name> (アナログメグオームハイテスタ<ANALOG MΩ HiTESTER>)
形名<Model Number> (IR4030) *1
製造番号<Serial No.> (No.180598765)
検査年月日<Test Date> (2018-05-09)
(<YYYY-MM-DD>)
検査条件<Test Conditions> (23.0 °C, 60 %rh)

項目 <Item>	レンジ <Range>	計器指示値 <Indicated Value>	許容範囲 <Tolerance>	校正値 <Calibration Value>
確度(許容差) <Accuracy (Limit Deviation Tolerance)>				
25V		0.005 MΩ	0.00450 MΩ ~ 0.00550 MΩ	(0.00500 MΩ)
		0.02 MΩ	0.0190 MΩ ~ 0.0210 MΩ	(0.0200 MΩ)
		0.5 MΩ	0.475 MΩ ~ 0.525 MΩ	(0.500 MΩ)
		10 MΩ	9.50 MΩ ~ 10.50 MΩ	(10.00 MΩ)
		20 MΩ	18.0 MΩ ~ 22.0 MΩ	(20.0 MΩ)
50V		0.005 MΩ	0.00450 MΩ ~ 0.00550 MΩ	(0.00500 MΩ)
		0.02 MΩ	0.0190 MΩ ~ 0.0210 MΩ	(0.0200 MΩ)
		0.5 MΩ	0.475 MΩ ~ 0.525 MΩ	(0.500 MΩ)
		10 MΩ	9.50 MΩ ~ 10.50 MΩ	(10.00 MΩ)
		20 MΩ	18.0 MΩ ~ 22.0 MΩ	(20.0 MΩ)
125V		0.005 MΩ	0.00450 MΩ ~ 0.00550 MΩ	(0.00500 MΩ)
		0.02 MΩ	0.0190 MΩ ~ 0.0210 MΩ	(0.0200 MΩ)
		0.5 MΩ	0.475 MΩ ~ 0.525 MΩ	(0.500 MΩ)
		10 MΩ	9.50 MΩ ~ 10.50 MΩ	(10.00 MΩ)
		20 MΩ	18.0 MΩ ~ 22.0 MΩ	(20.0 MΩ)
ACV (55Hz)	30 V	22.5 V	~ 37.5 V	(30.0 V)
	120 V	112.5 V	~ 127.5 V	(120.0 V)

備考<Note>

・ FAIL判定箇所は、グレー表示としています。<FAIL decision points are highlighted in gray.>

・ *1 形名欄の記載形名は、製品本体表記に基づいています。

<The model number stated in this document is based on the model number indicated on the instrument.>

総合判定<Overall Result>

(PASS)

検査者<Inspected By>

()

承認者<Approved By>

()